


ANTERIOR BRIDGES



IPS EMPRESS® 2

Note: Bridge connectors must be a minimum of 4.0 mm × 4.0 mm (16-mm sq.)





POSTERIOR CROWNS

IPS EMPRESS®, IPS EMPRESS® 2, IPS EMPRESS® ESTHETIC, IPS ERIS



POSTERIOR BRIDGES

IPS EMPRESS® 2



All prepared surfaces must be rounded on all metal-free restorations.